



# NATIONAL HEALTH MISSION

Government of Meghalaya

No. DHS/MCH&FW/NHM/RCH/65/2021/3311

Dated Shillong the 22.08.22

NOTICE

Following Technical bid opening which was held on 06<sup>th</sup>/June/2022 pertaining to Tender Document for "Supply and Installation Of Blood Bank Equipments, Instruments and Others" vide Tender Notice No: DHS/MCH&FW/NHM/RCH/65/2021/3236, dated 02nd/May/2022, placed below the findings and decision thereafter:

Sl. no.	Bidder's Name	Remarks
1	M/S Arti Enterprise, Elke Drugs & Distributor, M/S R.K.S Enterprise, Areng Medical Supplier R.G.S. Enterprise	Bidders are qualified
2	Thermofisher	Did not furnish Medical equipments customer feedback a. Medical equipments customer feedback b. A valid Trade License Certificate from KHADC/JHADC/ GHADC for Non-Tribal firm c. TECHNICAL BID: Annexure – II a. Bidder is disqualified

Sample evaluation:

Item no.	Name of the Item	Name of Bidders				
		Arti Enterprise	R.K.S. Enterprise	Areng Medical Supplier	Elke Drugs Distributor	R.G.S. Enterprise
1	Deep Freezer -40°C	Complied	Complied	<b>Not complied:</b> 1. Krish Biomedicals: Temperature requirement upto 40°C the range of this equipment is 20°C to 25°C 2. Acmas Technologies model ATI-127: Temperature chart recorder device not mentioned in catalogue. Detailed alarms 1-5 not mentioned 3. Icon Instruments: Catalogue does not specify about 1) Alarmer 1-5 2) Temperature chart recorder 4. Blue Star Limited: As per catalogue beep chart recorder facility not mentioned 5. Diamond Scientific and Surgicals: Catalogue does not specify whether it is upreplit model & other features required for plasma storage 6. Tempo Instruments Pvt. Ltd.: Temperature chart recorder details of & alarm not mentioned in catalogue <b>Complied:</b> 1. Acmas Technologies Model No: Weiber Deep Freezer 2. Middleby	Complied	Upright model required. -Details of A/V alarms not provided.
2	Blood Bank Refrigerator	Complied	Labworld Scientific Instruments: Details of temp chart recorder or alarms not mentioned in catalogue, other company are complied	<b>Not complied:</b> 1. Acmas Technologies Model No: ATI-200: Temperature chart recording facility not specified 2. Krish Biomedicals: Not as required as temperature range 3-16°C 3. Narang Medical Ltd: Alarms 1-5 not specified in catalogue 4. Diamond Scientific: Details of alarm not provided in catalogue <b>Complied:</b> 1. Acmas Technologies: Model No: Weiber Blood Bank refrigerator (ACM - 78095 S-I) 2. Middleby CelfrostInnovation	Complied	Should provide with details of Temperature chart recorder & A/V Alarms not mentioned

Draft #53 of File DHS/MCH&FW/NHM/RCH/65/2021 Approved by Mission Director on 18-Aug-2022 05:48 PM - Page 1

Office of Mission Director, National Health Mission

Directorate of Health Services, Health Complex, Upper New Colony, Laitumkhrah, Shillong - 793003

Phone: (0364) 2504532 Email: nrhmmegh@gmail.com

www.nrhmmeghalaya.nic.in

f Nhm Meghalaya



@iecbccnhmmegh



IECBCC NHM Meghalaya



# NATIONAL HEALTH MISSION

## Government of Meghalaya

Draft #53 of File DHS/MCH&FW/NHM/RCH/65/2021 Approved by Mission Director on 18-Aug-2022 05:48 PM - Page 2

3	Deep Freezer -80°C	Complied	Complied	<p><b>Not complied:</b></p> <ol style="list-style-type: none"> <li>Blue Star Limited: Catalogue does not mention temperature recording chart facility</li> <li>Krish Biomedicals: No Details of ink chart recorder in catalogue</li> <li>Diamond Scientific : As per catalogue model is horizontal &amp; not appropriate plasma storage</li> <li>Icon Instruments Company: Temperature range inadequate for plasma storage.</li> </ol> <p><b>Complied:</b></p> <ol style="list-style-type: none"> <li>Acmas Technologies Model No: ilabot &amp; Model No: ATJ-127</li> <li>Middleby CelfrostInnovation</li> </ol>	Complied	Storage capacity of 24-26 units required
4	Platelet Incubator	Complied	Complied	<p><b>Not complied:</b></p> <ol style="list-style-type: none"> <li>Narang Medical Ltd: Capacity not specified in catalogue in terms of units</li> <li>-Agitator not mentioned in catalogue</li> <li>-Alarms 1-5 not mentioned in catalogue</li> <li>Deluxe Scientific Model : BBPC 101 : capacity 48bags whereas request is 24-26 bags, detailed Alv Alarms 1-5 not specified</li> <li>Deluxe Scientific Model : BBPC 111 : Ink chart recorder not available</li> <li>Span Healthcare: Alarm &amp; sensor power failure not mentioned in catalogue</li> <li>Diamond Scientific : Details of alarm 1-5 not provided in catalogue</li> </ol> <p><b>Complied:</b></p> <ol style="list-style-type: none"> <li>Bioline India</li> <li>Middleby CelfrostInnovation</li> </ol>	Storage capacity for 24-26 bags required not 54 units	Complied
	Platelet Agitator	Complied	Labworld Scientific Instruments: Required capacity is 24-26 unit not 36/48 bags, other company are complied		Complied	
5	Semi Automated Coagulometer	Complied	<ol style="list-style-type: none"> <li>1 Channel</li> <li>Tests fibrinogen &amp; factors not mentioned</li> <li>Printer not mentioned</li> </ol>	Complied	Complied	Complied
6	Plasma Thwing Bath	Complied	Complied	<p><b>Not complied:</b></p> <ol style="list-style-type: none"> <li>Diamond Scientific and Surgicals: Temperature range not mentioned in catalogue</li> </ol> <p><b>Complied:</b></p> <ol style="list-style-type: none"> <li>Bioline India</li> <li>Narang Medical Ltd</li> <li>Span Healthcare Private Ltd</li> </ol>	Complied	Complied
7	Dielectric Tube Sealer	Complied	Complied	Complied	Complied	Complied
8	Blood Collection Monitor	Complied	Complied	Complied	Complied	Complied
9	Pre Vacuum Autoclave	Not Quoted	Not Quoted	Complied	Complied	Complied
10	Manual Plasma Expresser	Complied	Complied	Complied	Complied	Complied
11	Refrigerated Water Bath (Cryobath)	Complied	York Scientific Industries Pvt Ltd: Capacity not mentioned in terms of blood bag thawing, other company are complied	<p><b>Not complied:</b></p> <ol style="list-style-type: none"> <li>Tempo Instruments : Temperature range inadequate</li> <li>Diamond Scientific and Surgicals: Temperature range not in catalogue</li> </ol> <p><b>Complied:</b></p> <ol style="list-style-type: none"> <li>Acmas Technologies</li> <li>Narang Medical Ltd</li> <li>Span Healthcare Private Ltd</li> <li>Bioline India</li> <li>Acmas Technologies</li> </ol>	Complied	Capacity of 10-20 bags per run/1 cycle is required
12	Elisa Reader and Washer	Complied	Complied	Complied	Complied	Complied
	Elisa Reader and Washer	Complied	Complied	Complied	Complied	Complied
13	Table Top Centrifuge	Complied	Labworld Scientific Instruments & Cosmo Laboratory Equipments: 1. Rotor Head for 16 tubes to be procured at extra cost.	<p><b>Not complied:</b></p> <ol style="list-style-type: none"> <li>Deluxe Scientific: Capacity = 8/15 tubes and Requirement =16/24 tubes</li> </ol> <p><b>Complied:</b></p>	Complied	Complied

Office of Mission Director, National Health Mission

Directorate of Health Services, Health Complex, Upper New Colony, Laitumkhrah, Shillong - 793003

Phone: (0364) 2504532 Email: nrhmmegh@gmail.com

www.nrhmmehalaya.nic.in



Nhm Meghalaya



@iecbccnhmmegh



IECBCC NHM Meghalaya



# NATIONAL HEALTH MISSION

Government of Meghalaya

Draft #53 of File DHS/IMCH&FW/NHM/RCH/65/2021 Approved by Mission Director on 18-Aug-2022 05:48 PM - Page 3

			2.Product certification not available. other company are complied	1.Diamond Scientific		
14	Haemoglobinometer	Not Quoted	100 test strips to be provided not 10.	Complied	Not Quoted	Complied
15	Incubator	As per catalogue temperature range is only upto 60°C requirement 110°C	York Scientific Industries Pvt Ltd:Temperature range in adequate, other company are complied	<p><b>Not complied:</b></p> <p>1. Acmas Technologies Model No: ATI-117 : Temperature &amp; slot for thermometer not mentioned in catalogue</p> <p>2. Icon Instruments Company: Temperature range upto 70°C =In adequate</p> <p>3. Tempo Instruments : Temperature range not as per requirement</p> <p>4. Narang Medical Ltd: Catalogue does not mention</p> <p>1) Agitator facility for platelet storage</p> <p>2) Alarms 1-5</p> <p><b>Complied:</b></p> <p>1.Acmas Technologies Model No: (ACM - 22062-1 )</p> <p>2. Krish Biomedicals</p> <p>3. Span Healthcare Private Ltd</p> <p>3. Bioline India</p> <p>4. Acmas Technologies</p> <p>5. Diamond Scientific</p>	Temperature range only up to 60°C not 110°C as required	Complied
16	Water Bath	Complied	York Scientific Industries Pvt Ltd: Temperature range in adequate as it should be upto 100°C (as per catalogue), other company are complied	<p><b>Not complied:</b></p> <p>1. Acmas Technologies Model No: ATI-117 : Temperature &amp; slot for thermometer not mentioned in catalogue</p> <p>2. Tempo Instruments : Temperature range not as per requirement</p> <p>3. Deluxe Scientific : Temperature range is upto 80°C required range =22°C TO 100°C</p> <p>4. Diamond Scientific and Surgicals: Does not confirm to Specification temperature upto 90°C</p> <p><b>Complied:</b></p> <p>1.Acmas Technologies Model No: ATI-134 &amp; Weiber Serological Water Bath</p> <p>2. Icon Instruments Company</p> <p>3. Span Healthcare Private Ltd</p> <p>3. Bioline India</p> <p>4. Acmas Technologies</p> <p>5. Diamond Scientific</p>	Complied	Complied
17	Refrigerated Component Centrifuge	Maxspeed =4600 Requirement=10,000	Maxspeed 4200 whereas requirement is 10,000 speed	<p><b>Not complied:</b></p> <p>1. Acmas Technologies : Specifications in catalogue provided is for centrifuge &amp; not for preparation of blood components</p> <p>2. Narang Medical : As per catalogue speed is only 4,200rpm required speed is higher.</p> <p>3. Diamond Scientific : Centrifuge range is less, ie upto 4,500rpm</p>	Speed only upto 4,200 RPM required speed =10,000 rpm	NEW RED LAB EQUIPMENTS model no. RERC 700: Speed upto 10,000 rpm is required whereas this equipment has maxspeed of 4500/rpm. Other model are complied
18	Cell Counter (Automated Hematology Analyzer)	Complied	Complied	Complied	Complied	Not Quoted
			York Scientific Industries	<p><b>Not complied:</b></p> <p>1. Acmas Technologies Model No: ATI-128 : Should be horizontal flow model manometer to be provided</p> <p>2. Acmas Technologies Model No: ACM - 915 : Should be horizontal flow model manometer to be provided</p> <p>3. Tempo Instruments : Horizontal flow required</p>		Complied,

Office of Mission Director, National Health Mission

Directorate of Health Services, Health Complex, Upper New Colony, Laitumkhrah, Shillong - 793003

Phone: (0364) 2504532 Email: nrhmmegh@gmail.com

www.nrhmeghalaya.nic.in

f Nhm Meghalaya

@iecbccnhmmegh

IECBCC NHM Meghalaya



# NATIONAL HEALTH MISSION

Government of Meghalaya

Draft #53 of File DHS/MCH&FW/NHM/RCH/65/2021 Approved by Mission Director on 18-Aug-2022 05:48 PM - Page 4

19	Laminar Airtlow Bench (Bio-Safety Cabinet)	Complied	Pvt Ltd: Horizontal flow with manometer required, other company are complied	whereas output offered is vertical flow 4. Icon Instruments Company: Catalogue does not mention whether model is horizontal or not 5. Narang Medical Ltd: a.Monometer not provided b.Catalogue does not mention whether it is a horizontal flow model <b>Complied:</b> Deluxe Scientific Surgico Pvt Ltd	Not Quoted	however Should be provided with manometer
20	Donor Couch	Complied	Complied, however remote electronic face lit for head low position should be available	<b>Not complied:</b> 1. Narang Medical : Catalogue does not mention remote control for chart position for donor management <b>Complied:</b> 1. Diamond Scientific and Surgicals	Complied	Complied, however remote electronic face lit for head low position should be available
21	Sterile Connecting Device	Complied	Not Quoted	Complied	Not Quoted	Not Quoted
22	Micropipettes(Set)	Not Quoted	complied, however should be Multi channel pipettes not specified	Complied	Not Quoted	Complied
23	Micropipettes(Single Piece Adjustable)	Not Quoted	Complied	Complied	Not Quoted	Complied
24	Leukocyte removal filter for red cells	Complied	Not Quoted	Complied	Complied	Not Quoted
25	Hand Sealer with Tube roller & cutter	Complied	Not Quoted	Complied	Not Quoted	Not Quoted
26	Hot Air Oven	Not Quoted	Complied	Complied	Complied	Not Quoted
27	Sphygmonmanometer (Standing Type)	Not Quoted	Complied	Complied	Not Quoted	Not Quoted
28	Binocular Microscope	Complied	Complied	Complied	Not Quoted	Complied
29	Quality Control of ABO reagent (anti-A, anti-B, and anti-AB)	Complied	Not Quoted	Complied	Complied	Not Quoted
30	Acceptable quality of anti-globulin reagent	Complied	Complied	Not Quoted	Not Quoted	Not Quoted
31	Acceptable Titer and Avidity of ABO reagents	Complied	Not Quoted	Complied	Complied	Not Quoted
32	Acceptable Titer and Avidity of ABO reagents(continued)	Complied	Complied	Complied	Complied	Not Quoted
33	Quality Acceptable of Rh anti sera (Anti-D IgM & IgG, AntiC,c,E,e)	Complied	Not specified for IgM,IgG, AntiC,c,E,e	Complied	Complied	Not Quoted
34	Anti A1 Lectin	Complied	Complied	Complied	Complied	Not Quoted
	Anti-H Lectin	Complied	Not Quoted	Complied	Complied	Not Quoted
35	Bovine Serum Albumin (BSA)	Complied	Complied	Complied	Complied	Not Quoted
36	Blood Transportation Boxes	Not Quoted	Not Quoted	Complied	Not Quoted	Not Quoted
<b>List of Reagents</b>						
1	Elisa Test Kits for 4th gen Elisa Kit	Complied	Only detective of HCV antibodies whereas both abs & antigen detection is required	Complied	Complied	Not Quoted
2	HBSAG ELISA Kit	Complied	Complied	<b>Not complied:</b> 1.Medsorce: Detection limit 0.1ng/ml 2.Triviron Healthcare : Analytical sensitivity = 0.1ng/ml <b>Complied:</b> 1. Biogenix Inc Pvt. Ltd 2. Krishgen Biosystems 3. Oscar Medicare	Athenese-Dx: Requirement 06ng/ml Offer 0.1ng/ml, other company are complied	Not Quoted
				<b>Not complied:</b> 1.Lapseptems HIV 1+2 :Specification not as per		

Office of Mission Director, National Health Mission

Directorate of Health Services, Health Complex, Upper New Colony, Laitumkhrah, Shillong - 793003

Phone: (0364) 2504532 Email: nrhmmegh@gmail.com

www.nrhmeghalaya.nic.in

f Nhm Meghalaya



@iecbccnhmmegh



IECBCC NHM Meghalaya



# NATIONAL HEALTH MISSION

Government of Meghalaya

3	4th Gen HIV ELISA Kit	Complied	Complied	requirement as it does not detect HIV-Ag as per catalogue <b>Complied:</b> 1. Oscar Medicare Pvt. Ltd 2. Krishgen Biosystems 3. Oscar Medicare 4. Triviron Healthcare 5. 2. Biogenix Inc Pvt. Ltd.	Athenese-Dx: Requirement=1IU/ml Offer=5IU/ml , other company are complied	Not Quoted
4	Third generation ELISA for the detection of antibodies to Syphills	Complied	Complied	Complied	Complied	Not Quoted
5	Rapid Test for Malaria Pv/Pf	Complied	Complied	Complied	Complied	Complied

**Note:** The Technical bid opening will be held on 01<sup>st</sup>/Nov/2019 at 3:00pm in the office of the undersigned. All Tender committee members and qualified bidders only are requested to attend.

**M/S R.K.S Enterprise, Areng Medical Supplier Elke Drugs Distributor, Arti Enterprise and R.G.S Enterprise** are qualified for the Financial bid opening. However the price bid of only those items which were found technically complied and selected by the Technical Committee shall be considered for price bid evaluations and take into account.

**Note:** The Financial bid opening will be held on 30.08.2022 at 02:00 Pm in the office of the undersigned. All Tender committee members and qualified bidders are requested to attend.

Mission Director, NHM  
Meghalaya, Shillong.

Copy to:

1. Tender Committee members for information and necessary action

The document is digitally approved. Hence signature is not needed.